



**Dr. Jeyakumar Kandasamy**  
**Assistant Professor**

**DEPARTMENT OF CHEMISTRY**  
**INDIAN INSTITUTE OF TECHNOLOGY**  
**(BANARAS HINDU UNIVERSITY)**  
**VARANASI-221005**  
**UTTAR PRADESH, INDIA**  
**Cell: +91-9794516773**  
**Email: [jeyakumar.chy@iitbhu.ac.in](mailto:jeyakumar.chy@iitbhu.ac.in)**

IIT(BHU)/CHY/2018-19/JK/LC-MS-corrigendum-2

Dated: 12.10.2018

Corrigendum

Corrigendum for tender Id: IIT(BHU)/CHY/2018-19/JK/LC-MS/111, dated: 27-09-2018 related to the purchase of Liquid chromatography –Mass spectrometry.

**A: Modification is Specifications**

S. No.	Tender Terms/ Specifications	Modified Tender Terms/ Specifications
1.	<b>Mass Analyzer: Page No 19. S.No.3</b>	
1a	Mass axis stability should be $\pm 0.1$ amu over 12 hours.	Mass axis stability should be $\pm 0.1$ amu over 12 hours or 24 hours
2.	<b>PDA detector: Page 20. S.No.10</b>	
2a	Number of diode elements: 1024.	Number of diode elements 512 or 1024
2b	Data rate: 120 Hz or better.	Data rate: 100 Hz or better.

**B: Accessories**

10 KVA UPS with 60 mins backup for LC-MS may be quoted as Optional

**C: Bid Submission End Date**

Bid Submission End Date is extended till 01-11-2018 (06:00 PM)

**D: Bid Opening Date**

Bid Opening Date is 03-11-2018 (03:00 PM)

Dr. Jeyakumar Kandasamy (PI)

**Dr. JEYAKUMAR .K**  
Assistant Professor  
Department of Chemistry  
Indian Institute of Technology  
(Banaras Hindu University)  
Varanasi - 221 005  
Uttar Pradesh, India.