

## Corrigendum - I

For Tender Id: Che/18-19/T-RS/406 dated: 10-12-2018, for the purchase of **Portable UV Based COD, BOD, TOC, TSS, NO<sub>3</sub> and Detergent Analyzer**

The Annexure – 2 should be used as given below instead of given in tender document:-

### Annexure- 2

#### TECHNICAL COMPLIANCE STATEMENT

(To be submitted by bidder duly filled)

SL No.	Technical Requirement	YES/NO
<b>1.</b>	<b>Range (as mentioned or better)</b> Parameters                      CONCENTRATION RANGE (Mg/L) COD                                      1 – 350 BOD                                      0.2 - 350 TOC                                      0.5 - 300 NO <sub>3</sub> 0.5 - 40 NO <sub>3</sub> -N                                  0,11 – 9,03 TSS                                      2.5 - 100 Anionic detergent                  0.5 - 60	
<b>2.</b>	Accuracy and repeatability : +/-10% of measured value or better for all parameters	
<b>3.</b>	The instrument should measure all parameters simultaneously and should be interference free for all type of waste water	
<b>4.</b>	Response time should be less than 1 minute (Fast reading)	
<b>5.</b>	Direct measurement without use of reagent or consumable should be there in the system	
<b>6.</b>	Provision should be there in the instrument for results memorization, data traceability	
<b>7.</b>	Minimum warranty of 2 years should be included in the instrument	
<b>8.</b>	The company should provide training to the users for operation of the instrument	
<b>9.</b>	Availability of Service Canter in India	

Prof. R.S. Singh (PI)  
Deptt. of Chemical Engg. & Tech.  
IIT (BHU)