




भारतीय
प्रौद्योगिकी
संस्थान
काशी हिन्दू विश्वविद्यालय



INDIAN
INSTITUTE OF
TECHNOLOGY
BANARAS HINDU UNIVERSITY

भैषजकीय अभियांत्रिकी एवं
प्रौद्योगिकी विभाग

Department of Pharmaceutical
Engineering & Technology

No. IIT/PH/2018-19/477

Dated: 15.01.2019

CORRIGENDUM

Brief Description of Procurement: Pre-Clinical In- vivo Imaging System for Small
Animals: 1 complete Unit

Bid Ref.: IIT (BHU)/PH/FY/2018-19/QTN/M/58

Dated: 17.11.2018

This is in reference to the above tender; All the received technical bids are rejected in technical evaluation (technical evaluation report enclosed). Hence, as no any bid remained technically qualified for opening of financial bid so the above tender may be treated as cancelled.


(Prof. Sanjay Singh)
Head of the Department

भैषजकीय अभियांत्रिकी एवं प्रौद्योगिकी विभाग /
Department of Pharmaceutical Engineering & Technology
भारतीय प्रौद्योगिकी संस्थान / INDIAN INSTITUTE OF TECHNOLOGY
(बनारस हिन्दू विश्वविद्यालय) / (BANARAS HINDU UNIVERSITY)

Enclosure: Technical Evaluation report



भारतीय प्रौद्योगिकी संस्थान (का.हि.वि.) वाराणसी - 221005 भारत
Indian Institute of Technology (BHU) Varanasi-221005 India

Email: head.phe@itbhu.ac.in; Web: itbhu.ac.in



Technical Evaluation Report under E-procurement

Department : Department of Pharmaceutical Engg. & Tech., Indian Institute of Technology (Banaras Hindu University), Varanasi

Tender ID : 2018_IITBU_408904_1

Reference No. : IIT(BHU)/PH/FY/ 2018-19/QTN/M/58 Dated 17.11.2018

The above given tender that was published on the e-tender portal on dated 19-11-2018 has received the no. of bids 2 for Supply of Pre-Clinical In- vivo Imaging System for Small Animals: 1 complete Unit and both the following bidders are rejected.

Sl.No.	Bid ID	Bidder's Name	Instrument Name	Reason for Rejection
1	---	M/s. Fusion Scientific Laboratories Pvt. Ltd., Navi Mumbai, Maharashtra	Pre-Clinical In-vivo Imaging System for Small Animals: 1 complete Unit	Not all the supporting documents were provided in support of the technical specifications claimed.
2	---	M/s. Imperial Life Science Pvt Ltd, Gurgaon, Haryana.	Pre-Clinical In-vivo Imaging System for Small Animals: 1 complete Unit	Some of the important features were not complied such as real effective imaging pixel must be minimum 2030 x 2030 or better, etc.

Member of Technical Committee

- | | | |
|---------------------|---|--|
| 1. Prof. Sairam K. | - | DPC & In-vivo Technical Committee Chairman |
| 2. Dr. A.N. Sahu | - | In-vivo Technical Committee Member Secretary |
| 3. Dr. M.S. Muthu | - | In-vivo Technical Committee Member |
| 4. Dr. G.P. Modi | - | In-vivo Technical Committee Member |
| 5. Dr. S.K. Jain | - | In-vivo Technical Committee Member |
| 6. Dr. A.K. Agrawal | - | In-vivo Technical Committee Member |


Prof. Sairam K.


Dr. A.N. Sahu


Dr. M.S. Muthu


Dr. G.P. Modi


Dr. S.K. Jain


Dr. A.K. Agrawal